



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/660,110	09/11/2003	Mark F. Oldham	ABIOS.041A	6842

20995 7590 09/07/2006

KNOBBE MARTENS OLSON & BEAR LLP
2040 MAIN STREET
FOURTEENTH FLOOR
IRVINE, CA 92614

EXAMINER

NEGIN, RUSSELL SCOTT

ART UNIT	PAPER NUMBER
----------	--------------

1631

DATE MAILED: 09/07/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary	Application No. 10/660,110	Applicant(s) OLDHAM ET AL.	
	Examiner Russell S. Negin	Art Unit 1631	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 26 June 2006.
- 2a) ☐ This action is **FINAL**.
- 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 20-44 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 20-44 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☒ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☐ All b) ☐ Some * c) ☐ None of:
 - 1. ☐ Certified copies of the priority documents have been received.
 - 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 - 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☒ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date 1/12/2004.
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____.
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: _____.

DETAILED ACTION

Election/Restrictions

Applicant's election without traverse of Invention II (claims 1-19) in the reply filed on 26 June 2006 is acknowledged.

It is additionally acknowledged that the applicant cancels claims 1-19.

Specification

The disclosure is objected to because of the following informalities: In paragraph [0005], lines 6-7 state, "of a first particles" and lines 7-8 state, "of a second particles."

Appropriate correction is required.

Claim Rejections - 35 USC § 102

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 20-24 and 33-36 are rejected under 35 U.S.C. 102(b) as being anticipated by Morris [Nuclear Magnetic Resonance Imaging in Medicine and Biology, 1986, Oxford: Clarendon Press].

Claims 20-24 and 33-36 state:

20. A method for improving the measurement of one or more types of specific particles of a sample using a detector associated with a biological analysis system wherein the specific particles are adapted to emit identifiable signals based on the interaction of the specific particles with corresponding probes and wherein the identifiable signals are captured by the detector to yield an output signal and wherein the detector is adapted to be operated at different configurations that respond differently to the identifiable signals, the method comprising: performing a first measurement of the identifiable signals with the detector at a first configuration such that the detector yields a first output signal wherein the first configuration allows effective measurement of a first type of the specific particles; performing a second measurement of the identifiable signals with the detector

Art Unit: 1631

at a second configuration such that the detector yields a second output signal wherein the second configuration allows effective measurement of the second type of the specific particles; and combining the first and second output signals to obtain a representation of the identifiable signals wherein the representation of the identifiable signals includes effective representations of the first and second types of the specific particles to thereby allow improved identification of the specific particles within the sample.

21. The method of claim 20, wherein the first measurement at the first configuration is adapted to effectively measure a relatively strong component of the identifiable signals associated with the first type of the specific particles having a relatively high abundance.

22. The method of claim 21, wherein the second measurement at the second configuration is adapted to effectively measure a relatively weak component of the identifiable signals associated with the second type of the specific particles having a relatively low abundance.

23. The method of claim 22, wherein combining the first and second output signals comprises scaling the first output signal to a scale associated with the second configuration such that the based on the second configuration, the weak component is effectively measured and the strong component is effectively represented based on the scaling of the effectively measured value from the first configuration.

24. The method of claim 23, wherein the scaling of the strong component allows effective representation of both weak and strong components when a dynamic range associated with the detector is limited and would not be able to measure the strong component at the second configuration.

33. A method extending the effective dynamic range of a detector that measures detectable signals from a sample undergoing a biological analysis wherein the detectable signals comprise two or more components representative of two or more components of the sample, the method comprising: obtaining a first output signal from the detector operated at a first configuration that allows effective measurement of a first component of the detectable signals; obtaining a second output signal from the detector operated at a second configuration that allows effective measurement of a second component of the detectable signals wherein the second configuration is such that the first component of the detectable signals would fall outside the detector's dynamic range at the second configuration; and scaling the first output signal to a scale associated with the second configuration wherein the amount of scaling depends on the first and second configurations and wherein the scaled first output signal allows representation of the first output signal at the second configuration thereby extending the effective dynamic range of the detector and wherein such extension of the effective dynamic range allows improved characterization of the sample having a relatively large range of relative abundances of the two or more components.

34. The method of claim 33, wherein the first configuration is adapted to effectively measure a strong component of the detectable signals.

35. The method of claim 34, wherein the second configuration is adapted to effectively measure a weak component of the detectable signals.

36. The method of claim 35, wherein scaling the first output signal allows representation of both weak and strong components when the dynamic range associated with the detector is limited and would not be able to measure the strong component at the second configuration.

The technique that anticipates the pending set of claims is quadrature detection in Nuclear Magnetic Resonance.

Morris explains quadrature detection in section 2.4.2.3 of his book on pages 32-35. In spectrometer, quadrature detection is used for determining the direction of nuclear precession and to enhance signal to noise ratios. As explained on the first line of page 34, "In practice, one normally employs two p.s.d.s [phase sensitive detectors] with reference phases in quadrature (90° apart). This method is known as quadrature detection and is desirable since it allows a $\sqrt{2}$ improvement in S/N to be achieved..." The two phase sensitive detectors are synchronized to act as a single device used to transmit a single signal and spectrum.

Figure 2.11 on page 34 of Morris illustrates a signal detection device in a configuration along the x-axis used to deliver a first measurement of the identifiable signal. The initial measurement along the x-axis yields a high abundance as shown in Figure 2.11.

Figure 2.11 of page 34 of Morris illustrates a signal detection device in a configuration along the y axis use to deliver a second measurement of the identifiable

Art Unit: 1631

signal. The initial measurement along the y-axis yields a low abundance as shown in Figure 2.11.

The synchronization and combination of these signals yields a more strongly scaled single signal as a result of the dual method of detection with improved signal to noise ratios.

The types of particles measured are nuclei of atoms and are "nuclear spin systems" (page 32 of Morris, third line from the bottom).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

Claims 22-30 and 33-42 are rejected under 35 U.S.C. 103(a) as being unpatentable over Morris in view of Churchill et al [IEEE Transactions on Aerospace and Electronic Systems, Volume AES-17, no. 1, 1981] in view of Pierre et al [IEEE Transactions on Aerospace and Electronic Systems, 1995 pages 1900-1902] in view of Tacklind et al [US PGPUB 2003/0101605 A1].

20. A method for improving the measurement of one or more types of specific particles of a sample using a detector associated with a biological analysis system wherein the specific particles are adapted to emit identifiable signals based on the interaction of the specific particles with corresponding probes and wherein the identifiable signals are captured by the detector to yield an output signal and wherein the detector is adapted to be operated at different configurations that respond differently to the identifiable signals, the method comprising: performing a first measurement of the identifiable signals with the detector at a first configuration such that the detector yields a first output signal wherein the first configuration allows effective measurement of a first type of the specific particles; performing a second measurement of the identifiable signals with the detector at a second configuration such that the detector yields a second output signal wherein the second configuration allows effective measurement of the second type of the specific particles; and combining the first and second output signals to obtain a representation of the identifiable signals wherein the representation of the identifiable signals includes effective representations of the first and second types of the specific particles to thereby allow improved identification of the specific particles within the sample.

21. The method of claim 20, wherein the first measurement at the first configuration is adapted to effectively measure a relatively strong component of the identifiable signals associated with the first type of the specific particles having a relatively high abundance.

22. The method of claim 21, wherein the second measurement at the second configuration is adapted to effectively measure a relatively weak component of the identifiable signals associated with the second type of the specific particles having a relatively low abundance.

23. The method of claim 22, wherein combining the first and second output signals comprises scaling the first output signal to a scale associated with the second configuration such that based on the second configuration, the weak component is effectively measured and the strong component is effectively represented based on the scaling of the effectively measured value from the first configuration.

24. The method of claim 23, wherein the scaling of the strong component allows

Art Unit: 1631

effective representation of both weak and strong components when a dynamic range associated with the detector is limited and would not be able to measure the strong component at the second configuration.

25. The method of claim 24, wherein the detector is a charge-coupled device and the first configuration comprises a short exposure duration T_1 selected to effectively measure the strong component of the identifiable signals.

26. The method of claim 25, wherein the second configuration comprises a long exposure duration T_2 selected to effectively measure a weak component of the identifiable signals.

27. The method of claim 26, wherein the scaling of the first output signal comprises multiplying the first output signal value by a ratio T_2/T_1 .

28. The method of claim 24, wherein the detector is a charge multiplier and the first configuration comprises a low operating voltage V_1 selected to effectively measure the strong component of the identifiable signals.

29. The method of claim 28, wherein the second configuration comprises a high operating voltage V_2 selected to effectively measure a weak component of the identifiable signals.

30. The method of claim 29, wherein the scaling of the first output signal comprises determining the scaled value N_1' of the first output signal based on a relationship $\log(N_1') = m \log(V_2/V_1)$ where m represents a slope of a curve obtained by plotting the multiplier's gain versus the voltage in a log-log manner.

33. A method extending the effective dynamic range of a detector that measures detectable signals from a sample undergoing a biological analysis wherein the detectable signals comprise two or more components representative of two or more components of the sample, the method comprising: obtaining a first output signal from the detector operated at a first configuration that allows effective measurement of a first component of the detectable signals; obtaining a second output signal from the detector operated at a second configuration that allows effective measurement of a second component of the detectable signals wherein the second configuration is such that the first component of the detectable signals would fall outside the detector's dynamic range at the second configuration; and scaling the first output signal to a scale associated with the second configuration wherein the amount of scaling depends on the first and second configurations and wherein the scaled first output signal allows representation of the first output signal at the second configuration thereby extending the effective dynamic range of the detector and wherein such extension of the effective dynamic range allows improved characterization of the sample having a relatively large range of relative abundances of the two or more components.

Art Unit: 1631

34. The method of claim 33, wherein the first configuration is adapted to effectively measure a strong component of the detectable signals.

35. The method of claim 34, wherein the second configuration is adapted to effectively measure a weak component of the detectable signals.

36. The method of claim 35, wherein scaling the first output signal allows representation of both weak and strong components when the dynamic range associated with the detector is limited and would not be able to measure the strong component at the second configuration.

37. The method of claim 36, wherein the detector is a charge-coupled device and the first configuration comprises a short exposure duration T_1 selected to effectively measure the strong component of the detectable signals.

38. The method of claim 37, wherein the second configuration comprises a long exposure duration T_2 selected to effectively measure a weak component of the detectable signals.

39. The method of claim 38, wherein the scaling of the first output signal comprises multiplying the first output signal value by a ratio T_2/T_1 .

40. The method of claim 36, wherein the detector is a charge multiplier and the first configuration comprises a low operating voltage V_1 selected to effectively measure the strong component of the detectable signals.

41. The method of claim 40, wherein the second configuration comprises a high operating voltage V_2 selected to effectively measure a weak component of the detectable signals.

42. The method of claim 41, wherein the scaling of the first output signal comprises determining the scaled value N_1' of the first output signal based on a relationship $\log(N_1') = m \log(V_2/V_1)$ where m represents a slope of a curve obtained by plotting the multiplier's gain versus the voltage in a log-log manner.

The technique that anticipates the pending set of claims is quadrature detection in Nuclear Magnetic Resonance.

Morris explains quadrature detection in section 2.4.2.3 of his book on pages 32-

35. In spectrometer, quadrature detection is used for determining the direction of

Art Unit: 1631

nuclear precision and to enhance signal to noise ratios. As explained on the first line of page 34, "In practice, one normally employs two p.s.d.s [phase sensitive detectors] with reference phases in quadrature (90° apart). This method is known as quadrature detection and is desirable since it allows a $\sqrt{2}$ improvement in S/N to be achieved..."

The two phase sensitive detectors are synchronized to act as a single device used to transmit a single signal and spectrum.

Figure 2.11 on page 34 of Morris illustrates a signal detection device in a configuration along the x-axis used to deliver a first measurement of the identifiable signal. The initial measurement along the x-axis yields a high abundance as shown in Figure 2.11.

Figure 2.11 of page 34 of Morris illustrates a signal detection device in a configuration along the y axis use to deliver a second measurement of the identifiable signal. The initial measurement along the y-axis yields a low abundance as shown in Figure 2.11.

The synchronization and combination of these signals yields a more strongly scaled single signal as a result of the dual method of detection with improved signal to noise ratios.

Morris further illustrates two different sets of times on the coordinate systems of Figure 2.11- a coordinate system with a time for the x-axis detector configuration and a different coordinate system with a time for the y-axis detector configuration system.

The types of particles measured are nuclei of atoms and are "nuclear spin systems" (page 32 of Morris, third line from the bottom).

Morris fails to take ratios of these times, show specific voltages employed to regulate the quadrature detector, show the required logarithmic plots, and utilize a charge-coupled device.

Churchill et al states in the article, "The Correction of I and Q Errors in a Coherent Processor," "A method is presented for correcting the gain and phase imbalances and the bias errors of the in-phase and quadrature channels of a coherent signal processor... The residual errors after correction depend upon the signal-to-noise ratio (S/N) of the test signal and the degree of filtering used in deriving the correction coefficients." The image powers used in the correction of the image power (amplitude and phase) imbalances are shown in the middle of the second column of page 131 and apply to both channels.

However, Churchill et al fails to manipulate the power values in a method consistent with the requirements of the instant claims or teach the use of a CCD device.

In the article of Pierre et al, entitled, "Considerations in the autocalibration of quadrature receivers," Figure 2 on page 1902 shows a log-log plot of the variance of the signal to noise ratio as a function of the ratio of two times. The plotted data simulates a line with a specific slope. It would be an obvious mathematical transformation to plot voltage or power ratios as determined in Churchill et al in place of the ratio of times found in Pierre et al.

However none of the above sources show use of charge coupled devices in signal detection.

Tacklind et al, in their invention entitled, "Servo-controlled automatic level and plumb tool," explain that quadrature detectors and charge coupled devices are interchangeable (paragraph [0060]): "The position sensitive photo sensor 131 can incorporate any of a number of commercially available position sensitive detectors sensitive to the detector light 134. Examples include, but are not limited to, quadrature detectors, charged coupled devices (CCD) detectors, complementary metal oxide semiconductor (CMOS) image sensors..."

It would have been obvious to someone of ordinary skill in the art at the time of the instant invention to modify Morris in view of the voltage determination method of Churchill et al in view of the autocalibration method of Pierre et al in view of the CCD substitution method of Tacklind et al because while Morris teaches how quadrature detection applies to NMR, Churchill et al teaches how image power could be used for correction of imbalances in quadrature detection, Pierre et al teaches how log-log plots correct for considerations in the autocalibration of quadrature receivers, and Tacklind et al teach how charge coupled devices and replaceable by quadrature detectors for an analogous application.

Claims 20-24, 28-36, and 40-44 are rejected under 35 U.S.C. 103(a) as being unpatentable over Morris in view of Churchill et al in view of Pierre et al in view of Tacklind et al as applied to claims 22-30 and 33-42 in further view of Photomultiplier Tubes [Hamamatsu Brochure, pages 1-15, July 2002].

Claims 31-32 and 43-44 state:

31. The method of claim 30, wherein the charge multiplier comprises a photomultiplier tube.

32. The method of claim 30, wherein the charge multiplier comprises a charge intensifier.

43. The method of claim 42, wherein the charge multiplier comprises a photomultiplier tube.

44. The method of claim 42, wherein the charge multiplier comprises a charge intensifier.

While the previously explained sources discuss the method steps of the instant set of claims, they do not disclose the photomultiplier tube or the charge intensifier.

The catalog "Photomultiplier tubes" discusses uses and sales of photomultiplier tubes and charge intensifiers throughout the brochure.

It would have been obvious at the time of the instant invention for someone of ordinary skill in the art to modify Morris in view of Churchill et al in view of Pierre et al in view of Tacklind et al as applied to claims 22-30 and 33-42 in further view of Photomultiplier to result in the instant invention because Photomultiplier has the advantage of selling the relevant items necessary to improve the claimed invention.

Conclusion

No claim is allowed.

Papers related to this application may be submitted to Technical Center 1600 by facsimile transmission. Papers should be faxed to Technical Center 1600 via the central PTO Fax Center. The faxing of such pages must conform with the notices published in the Official Gazette, 1096 OG 30 (November 15, 1988), 1156 OG 61 (November 16, 1993), and 1157 OG 94 (December 28, 1993)(See 37 CFR § 1.6(d)). The Central PTO Fax Center Number is (571) 273-8300.

Art Unit: 1631

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Russell Negin, Ph.D., whose telephone number is (571) 272-1083. The examiner can normally be reached on Monday-Friday from 7am to 4pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's Supervisor, Andrew Wang, Supervisory Patent Examiner, can be reached at (571) 272-0811.

Any inquiry of a general nature or relating to the status of this application should be directed to Legal Instrument Examiner, Yolanda Chadwick, whose telephone number is (571) 272-0514.


Information regarding the status of the application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information on the PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

-RSN

5 September 2006



5 Sept 2006

 5 September 2006
JOHN S. BRUSCA, PH.D.
PRIMARY EXAMINER